



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Precise Dimensional Inspection LLC
250 West Temperance Road, Temperance, MI 48182

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Initial Accreditation Date:

April 17, 2020

Issue Date:

October 18, 2022

Expiration Date:

December 31, 2024

Tracy Szerszen
President

Accreditation No.:

109839

Certificate No.:

L22-700

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Precise Dimensional Inspection LLC

250 West Temperance Road, Temperance, MI 48182

Contact Name: Jacob Bieniek Phone: 734-847-6858

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calipers ^F	Up to 12 in Up to 24 in	(150 + 25L) μ in (220 + 28L) μ in	Gladstonbury Standard Gage Blocks Size control standard PRO-0103
Outside Micrometers ^F	Up to 12 in	(43 + 10L) μ in	Gladstonbury Master PRO- 0102
Plug Gages ^F	Up to 6 in	28 μ in	Mahr ULM 828 PRO-0124
Ring Gages ^F	Up to 12 in	28 μ in	Mahr ULM 828 PRO- 0123
Check Fixtures ^F	X Axis 16 in Y Axis 16 in Z Axis 16 in	(100 + 5.2L) μ in	CMM PRO-0117
Flush Pin Gages ^F	Up to 8 in	(100 + 5.2L) μ in	CMM PRO-0120
Reference Spheres ^F	Up to 2 in	(100 + 5.2L) μ in	CMM PRO-0115
Step Gages ^F	Up to 12 in	(100 + 5.2L) μ in	CMM PRO-0116
Thread Gages ^F	4 to 40 TPI in	28 μ in	Mahr ULM 828 PRO-0125 3 wire Method
Surface Finish ^F	Up to 4mm	4.3 μ in	Mahr Surf PS10 PRO-0126
General Length Measurement	Up to 40 inches	28 μ in	Mahr Model #828 PRO-0122

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.



Certificate of Accreditation: Supplement

Precise Dimensional Inspection LLC

250 West Temperance Road, Temperance, MI 48182

Contact Name: Jacob Bieniek Phone: 734-847-6858

Accreditation is granted to the facility to perform the following calibrations:

3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
5. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
6. The above scope of accreditation was created based on a former ILAC MRA Signatory's certificate policy. Based on the intent of the ILAC MRA, PJLA recognizes other scopes issued by other ILAC signatories. This scope will be modified based on PJLA's Policy following the next on-site assessment.

